



Certificate of Traceability

Manufacturer Certification

This certifies that **Tosoh Corporation**, using **2.2 min Variant Analysis Mode on G7** has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#9**

The system evaluated was:

| | | |
|--|---|-------------------------|
| Instrument: G7 | Calibrator Lot: ZS6002 | Column Lot: M |
| Reagent Lot: M7-109A, M7-204A, M7-304A, HW-20B | Calibrator Assigned Values: 5.53%, 10.63% | |

Date of Certification: June 1, 2017

Certification Expires: June 1, 2018

David Sacks

NGSP Steering Committee Chair

Randie R Little PhD.

NGSP Network Coordinator

Shawn Amalfi

SRL director/ supervisor