



Certificate of Traceability

Manufacturer Certification

This certifies that **Tosoh Corporation**, using **1.6 min Variant Analysis Mode on G8** has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#9**

The system evaluated was:

Instrument: G8	Calibrator Lot: ZS6002	Column Lot: H
Reagent Lot: H8-101A, H8-201A, H8-301A, HW-20B	Calibrator Assigned Values: 5.53%, 10.63%	

Date of Certification: June 1, 2017

Certification Expires: June 1, 2018

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor