



# Certificate of Traceability

## Manufacturer Certification

This certifies that **Tosoh Corporation**, using **1.0 min Variant Analysis Mode on G11** has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#9**

The system evaluated was:

Instrument: <b>G11</b>	Calibrator Lot: <b>ZS9002</b>	Column Lot #: <b>B</b>
Reagent Lot: <b>B1-101F, B1-201F, B1-306E, HW-30E</b>	Calibrator Assigned Values: <b>5.84%, 10.71%</b>	<b>Ver.3.06</b>

Date of Certification: June 1, 2020

Certification Expires: June 1, 2021

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor