



Certificate of Traceability

Manufacturer Certification

This certifies that *Tosoh Corporation*, using *G8 1.6 min Variant Analysis Mode Ver.5.24F* has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#9**

The system evaluated was:

Instrument: HLC-723G8	Calibrator Lot: ZS2002	Column Lot: S
Reagent Lot: S8-106H, S8-208H, S8-305H, HW-44H	Calibrator Assigned Values: 5.84%, 10.94%	

Date of Certification: June 1, 2023

Certification Expires: June 1, 2024

A handwritten signature in black ink, appearing to read 'David Sacks', is written over a horizontal line.

NGSP Steering Committee Chair

A handwritten signature in black ink, appearing to read 'Randie R Little PhD', is written over a horizontal line.

NGSP Network Coordinator

A handwritten signature in black ink, appearing to read 'Sharon Anulf', is written over a horizontal line.

SRL director/ supervisor